

### nitrylex® classic



PRODUCT DESCRIPTION		
Manufacturer	MERCATOR MEDICAL S.A. Ul. H. Modrzejewskiej 30 31-327 Kraków	
Type of the glove / Intended use	Non-sterile, powder-free, examination and protective glove for single use	
Material	Nitrile	
Donning powder	None	
Colour	Blue	
Shape	Ambidextrous, gloves fitting either hand	
Cuff	Beaded	
External surface	Microtextured + fingertip textured	
Internal surface	Chlorinated	
Packaging	10 x 100 pcs	10 x 200 pcs

PRODUCT REFERENCES			
Size / reference number	XS	RD30019001	RD30096001
	S	RD30019002	RD30096002
	M	RD30019003	RD30096003
	L	RD30019004	RD30096004
	XL	RD30019005	RD30096005

PHYSICAL PROPERTIES							
Size			XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)
Length [mm]	Spec. [min]		240	240	240	240	240
Width [mm]	Spec.		≤80	80 ±10	95 ±10	110 ±10	≥110
Thickness (single wall) [mm]	Middle finger	Spec. [min]	0,06				
	Palm	Spec. [min]	0,05				
	Cuff	Spec. [min]	0,04				
Elongation at break [%]	Before ageing	Spec. [min]	500				
	After ageing	Spec. [min]	400				
Force at break [N]	Before ageing	Spec. [min]	6,0				
	After ageing	Spec. [min]	6,0				
Powder content [mg/glove]	Spec.		<2				
Latex protein content [µg/g]	Spec.		Not applicable				

MANUFACTURING AND SAFETY STANDARDS		
AQL	Manufacturing final release: G-I inspection level AQL 1.0 in accordance with ISO 2859-1	
Latex protein content	Not applicable. Gloves do not contain natural rubber latex.	
Residual powder content	<2 mg/glove (according to EN 455-3)	
Classification	<b>Medical Device: CE</b> Class I (Regulation (UE) 2017/745)	<b>Personal Protective Equipment: CE 2777</b> Category III (Regulation (UE) 2016/425) Type B (EN ISO 374-1)
Product compliances	EN 455-1, EN 455-2, EN 455-3, EN 455-4, EN ISO 15223-1, EN 1041	EN ISO 374-1, EN ISO 374-2, EN 374-4, EN ISO 374-5, EN 16523-1, EN 420
Quality compliances	EN ISO 13485, ISO 9001	
Viral test permeation	Test in accordance with ASTM F1671 and EN ISO 374-5 (ISO 16604)	
Chemotherapy drugs permeation test	Test in accordance with ASTM D6978.	
Chemical substances permeation test	Test in accordance with EN 16523-1	
Biocompatibility / biological evaluation	Test in accordance with EN ISO 10993-5. No cytotoxic evidence observed. Test in accordance with EN ISO 10993-10. No skin irritation and sensitization evidence observed.	
Food contact	Gloves intended for food contact based on Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling any type of food and have been tested for Overall Migration Test acc. EN 1186.	
STORAGE AND DISPOSAL		
Storage instruction	It is recommended to store the gloves in dry place, in the temperature of 5-35°C and to protect them against direct sunlight. Keep the gloves in a distance of not less than 1m from heating devices, sources of fire and ozone. Do not keep in direct vicinity of solvents, oils, fuels and lubricants.	
Shelf life	3 years	
Disposal	Used gloves should be treated as a contaminated material, therefore local regulations regarding the disposal of such materials should be applied.	